



# Certificate of Calibration

**Certificate No:** D111111B0D01  
**Manufacturer:** JSB TECH Pte Ltd  
**Model No:** DWL3500XY  
**Serial No:** S/N 11A11111

**Date:** 1<sup>st</sup> January 2018  
**Description:** Dual Axis Precision Digital Level & Vibrometer

**Options Installed with Specifications:** Temperature Compensated to -10° C to + 50° C

**Date of Calibration:** 1<sup>st</sup> January 2018 (in storage: next cal. 5 years)

**Temperature:** 20° C ± 1° C

**Humidity:** 55 ± 10 %RH

**Procedure:** DP\_3500XYV1V1/temp 382793093

This certifies that the above product (s) was calibrated in compliance with ISO9001:2015 using applicable Digi-Pas® procedures. These calibration procedure and test points are those recommended in the procedure developed by Digi-Pas®.

**As received:** Factory tested

**As Shipping Conditions:** At the completion of the calibration, measure values were In-Specification at the points tested.

**Remarks or Special Requirements:** The product (s) bear a CE and/or FCC marks which is tested and certified by TÜV SÜD & PSB to comply with the EC and/or USA directives.

**Traceability Information:** Supporting document related to traceability is available for review by appointment. This report shall not be reproduced except in full, without prior written approval of the calibration facility.

**Calibration Equipment Used:**

<u>Model Number</u>	<u>Model Description</u>
SH-240 S3	ESPEC -TAIBAI Temperature & Humidity Chamber
P3- KPI897803	Digi-Pas® Reliability, Test & Evaluation Equipment
GT-0401008	Mitutoyo Granite Table (Grade Zero)
MS-308405	Starrett Master Square (Grade AA)
SB-283809	Sine Bar
BG-516-946R-10	Mitutoyo Block Gauge (Grade Zero)
LL000076 & LL001230	Digi-Pas® Precision Angle Jigs

**Test/Inspection Report:**

<u>Test Description</u>	<u>Specifications</u>	<u>Results</u>	<u>Judgments</u>
Resolution:	0.001° (≤20µm/M)	OK	PASS
Accuracy Cal:	±0.001° (0.000° ~ ±2.000°) ±0.003° for other angles	±0.001° ±0.003°	PASS PASS
Repeatability:	±0.001°	±0.001°	PASS
Symmetrical Check:	±0.001°	±0.001°	PASS
Metal Base Surface Flatness:			
Base Surface:	8µm	6µm	PASS
Vertical Surface:	8µm	7µm	PASS
Metal Base Angle:	90°± 0.01°	90°± 0.01°	PASS
Buzzer (Smart sensor):	>50dBA	64dBA	PASS
LCD display:	TFT LCD	OK	PASS
Operating Temp.:	-10° C to + 50° C	-15° C to + 55° C	PASS
Storage Temp.:	-20° C to + 70° C	-25° C to + 75° C	PASS

The result of this Certificate of Calibration apply only to the item described above. All reported accuracy values are as measured with no reduction by the uncertainty of measurement. Measurement uncertainty is expresses at a confidence level of 95% (coverage factor k=2).



Certificate No : D111111B0D01  
Serial Number : S/N 11A11111

Uncertainty: ±0.002  
Units : Degrees

**Single Axis:**


Description	Standard	Lower Limit	Upper Limit	As Found
<b>(+) SLOPE</b>	0.000	-0.001	0.001	0.000
	1.003	1.002	1.004	1.004
	1.433	1.432	1.434	1.432
	2.006	2.003	2.009	2.009
	2.866	2.863	2.869	2.866
	5.020	5.017	5.023	5.023
	10.079	10.076	10.082	10.076
	15.070	15.067	15.073	15.073
	19.877	19.874	19.880	19.875
<b>(-) SLOPE</b>	0.000	-0.001	0.001	0.000
	1.003	1.002	1.004	1.004
	1.433	1.432	1.434	1.432
	2.006	2.003	2.009	2.008
	2.866	2.863	2.869	2.867
	5.020	5.017	5.023	5.022
	10.079	10.076	10.082	10.076
	15.070	15.067	15.073	15.073
	19.877	19.874	19.880	19.876

**Dual Axis:**

Description	Standard	Lower Limit	Upper Limit	As Found	
<b>X-axis</b>					
<b>(+) SLOPE</b>	0.000	-0.001	0.001	0.000	
	1.003	1.002	1.004	1.004	
	1.433	1.432	1.434	1.434	
	2.006	2.003	2.009	2.003	
	2.866	2.863	2.869	2.869	
	5.020	5.017	5.023	5.023	
	9.933	9.930	9.936	9.934	
	<b>(-) SLOPE</b>	0.000	-0.001	0.001	0.000
		1.003	1.002	1.004	1.004
1.433		1.432	1.434	1.434	
2.006		2.003	2.009	2.004	
2.866		2.863	2.869	2.869	
5.020		5.017	5.023	5.022	
9.933		9.930	9.936	9.935	
<b>Y-axis</b>					
<b>(+) SLOPE</b>		0.000	-0.001	0.001	0.000
	1.003	1.002	1.004	1.004	
	1.433	1.432	1.434	1.433	
	2.006	2.003	2.009	2.003	
	2.866	2.863	2.869	2.865	
	5.020	5.017	5.023	5.018	
	9.933	9.930	9.936	9.933	
	<b>(-) SLOPE</b>	0.000	-0.001	0.001	0.000
		1.003	1.002	1.004	1.004
1.433		1.432	1.434	1.433	
2.006		2.003	2.009	2.004	
2.866		2.863	2.869	2.865	
5.020		5.017	5.023	5.017	
9.933		9.930	9.936	9.932	

Checked By: \_\_\_\_\_

Wu Ping  
QA Engineer



Laboratory Authorized: \_\_\_\_\_

Kuu  
Manager

